



# URGENT MEDICAL DEVICE RECALL

## RG554H2 SPACE SAVER SHOWER STOOL



### Do you own this product?

REDGUM Brand Space Saver Shower Stool

RG554H2 – 4-Legged Shower Stool

Dates sold: 6/10/2022 – 12/12/2023



Why the product is recalled: Cracking Seat Area

Hazard: Seat section failure causing injury



1. STOP using this product immediately
2. Contact the retailer you purchased the product from for a full refund
3. Remove the legs & handles from the unit
4. Dispose of the unit in general waste
5. Send us an email or phone us if you are unable to locate the product or have sold or gifted it to another person



Ask our staff to find out more about the recall.

[sales@redgumbrand.com.au](mailto:sales@redgumbrand.com.au)

1800 672 633 8.00am to 4.00pm Monday to Friday



Purchased from: Major Mobility Equipment Retailers / Online Sellers



Scan for information about this recall or go to  
<https://www.redgumbrand.com.au/>



Set yourself a reminder to check at home

- Take a photo of this notice
- Set a reminder on your phone

Further details regarding this recall can be found on the following website  
<https://www.tga.gov.au/>